



World Precision Instruments, Inc.
175 Sarasota Center Boulevard
Sarasota, FL 34240



ELI Accession Number: WPI-7981-1217

Date of completion: 12-15-2017

Lot number: 26102017

Reference number: N/A

Description of test article: FluoroDish™

Assay system requested by customer: Two cell mouse embryos were cultured in the test article using three 50µl drops of culture medium and overlaid with oil for 72-hours.

Control assay method and results: 15 two cell (B6D2F1 X B6C3F1) embryos were cultured in triplicate micro drops of culture medium and overlaid with oil:

15 / 15 (100 %)

2-cell to expanded blastocyst within 72 hr

For a valid assay, Embryotech™ requires at least 80% of two cell stage control embryos to develop to expanded blastocyst within 72-hours.

Test assay method and results: 21 two cell (B6D2F1 X B6C3F1) embryos were cultured in the test article using culture medium and overlaid with oil:

21 / 21 (100 %)

2-cell to expanded blastocyst within 72 hr

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos and were cultured in the same incubator at 37°C and 5.0% CO₂. 100 percent of the control embryos developed to the expanded blastocyst stage within 72-hours. 100 percent of the test embryos cultured in the test article developed to the expanded blastocyst stage within 72-hours

Signature
Study Director

12-15-2017
Date

Signature
Quality Reviewer

12-15-2017
Date